

ISTRUZIONI D'USO

Art. AK102

SUPPORTO LOMBARE

Grazie per aver scelto NewellMed, sinonimo di qualità e competenza nel settore ortopedico. Leggere attentamente le istruzioni seguenti per un corretto utilizzo del presidio.

INDICAZIONI TERAPEUTICHE

indicato nella cura /prevenzione lombalgie e tutte le affezioni del tratto lombare

PRECAUZIONI D'USO

-Posizionare il tutore senza esercitare pressioni eccessive ed evitando l'applicazione su parti che presentino ferite o eccessivi gonfiori.

-Per ogni dubbio applicativo rivolgersi al Tecnico Ortopedico di fiducia.

-Si consiglia di indossare il dispositivo lontano da fiamme libere.

- Si consiglia la sostituzione del dispositivo quando si presenta usura o logorio anche parziale.

MANUTENZIONE

-Lavaggio a mano con temperature max 30° utilizzando sapone neutro.

-Non candeggiare

-Non stirare

-Asciugare lontano da fonti di calore e conservare in luoghi freschi



APPLICAZIONE

- 1) Tensionare il busto anteriormente facendolo ben aderire al corpo.
- 2) Regolare la spinta lombare mediante i tiranti laterali che possono essere regolati sfruttando il tessuto asolato del presidio.

NOTE ED AVVERTENZE

Dispositivo medico di Classe I, fabbricato dalla Contel S.r.l. nella sua sede Strada delle Brigole, 46 23877 paderno d'Adda (LC) Italia in conformità alla Direttiva 93/42 CE.

Per l'utilizzo di questo dispositivo è consigliabile consultare un tecnico ortopedico od il medico curante per evitare ogni contro indicazione. Il giusto funzionamento del tutore e conseguente ad una corretta applicazione. Si consiglia l'utilizzo personale del prodotto. In caso di malfunzionamento rivolgersi al proprio riveditore di fiducia o contattare direttamente Contel S.r.l.. L'azienda non risponde in caso di uso improprio

USER GUIDE

Art. AK102

LUMBAR SUPPORT

Thank you for choosing NewellMed synonym of quality and competence in orthopedic field. Please read the following instruction carefully for a correct use of the product.

THERAPEUTIC INDICATIONS

Care an prevention of lumbar back pain, in all lumbar injuries

PRECAUTIONS FOR USE

-Position the support without exerting too much pressure and avoiding the application on injured or excessive swelling.

-For any question, please contact the Orthopaedic Technician.

-It's not raccomandated to wear the support near open flames

-It is raccomandated to replace the device when there is wear or tear, even partial

MAINTENANCE

-hand washing with maximum temperature 30° using mild soap.

-Don't bleach.

-Don't iron.

-Dry away from direct heat sources and sotre in a cool place.



APPLICATION INSTRUCTION

- 1) Front fastening. The support should be adherent to the body
- 2) Adjust the back push through lateral tie-rods

NOTES AND WARNINGS

Class I medical device, manufactured by Contel S.r.l. at its headquarters in Strada delle Brigole, 46 23877 Paderno d'Adda (LC) Italy in accordance with Directive 93/42 CE.

For the use of this device is advisable to consult a doctor or orthopedic technician to avoid any contraindications.

The proper functioning of the brace is the result of a correct application We reccomend personal use of the product.

In case of malfunction contact your local dealer or contact Contel S.r.l.. The company is not liable for misure of the product.